

## ***Instructions for Valve Replacement***

1. Immerse the old valve in hot water for 1-2 minutes, taking care not to get the cushion wet.
2. With pliers, work the old valve from side to side and then pull straight out of the cushion.
3. Using the Repair Adhesive cap, pierce seal in tube.
4. Squeeze a thin bead of adhesive around the barbed end of the new valve.
5. Press and twist the valve into the cushion until it will go no further. Remove any excess adhesive.
6. Allow to cure for three hours. Your cushion is now ready to use.

***Having trouble with your repair?***

***Contact Customer Service for assistance or an RA number:  
(800) 827-4548***



©2011 VARILITE  
4000 First Avenue South, Seattle, WA 98134 U.S.A.  
1-206-505-9500 | 1-800-827-4548  
www.varilite.com  
Made in the U.S.A